



Certificate of Compliance

Certificate: 2699955

Master Contract: 225290

Project: 80113731

Date Issued: February 28, 2022

Issued To: Eppendorf Himac Technologies Co., Ltd.
Takeda 1060
Hitachinaka-shi, Ibaraki, 312-8502
Japan
Attention: Kazuhiko Murayama

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.




Issued by: Massimo De Melis

PRODUCTS

CLASS - C8721 86 - ELECTRICAL LABORATORY EQUIPMENT Certified to US Standards

CLASS - C8721 06 - ELECTRICAL LABORATORY EQUIPMENT Electrical Laboratory Equipment

Ultracentrifuge, floor standing, permanently or cord connected, grounded, models CP100NX, CP90NX and CP80NX, rating: 208, 220, 230 or 240 V, 50/60 Hz, 20, 20, 16 or 16 A, Single phase.

Notes:

1. The above model is cord connected, Equipment Class I, Pollution Degree 2.
2. Mode of operation: Continuous.
3. Environmental Conditions: Extended: 2 to 40°C, 2000 m max, 80% to temperatures up to 31 °C decreasing linearly to 50% rH at 40 °C.



Certificate: 2699955
Project: 80113731

Master Contract: 225290
Date Issued: February 28, 2022

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA C22.2 No. 61010-1-12, UPD1: 2015,
UPD2: 2016, AMD1: 2018

CAN/CSA C22.2 No. 61010-2-020:17

Safety Requirements for Electrical Equipment for
Measurement, Control and Laboratory Use, Part 1:
General Requirements

Safety Requirements for Electrical Equipment For
Measurement, Control and Laboratory Use, Part 2-020:
Particular Requirements for Laboratory Centrifuges

UL Standards:

UL 61010-1, 3rd edition (2012), AMD1: 2018

UL 61010-2-020 3rd Ed.

Electrical Measurement and Test Equipment for
Measurement Use, Part I: General Requirements
Safety Requirements for Electrical Equipment For
Measurement, Control and Laboratory Use, Part 2-020:
Particular Requirements for Laboratory Centrifuges

Conditions of Acceptability:

- 1) The equipment is supplied with an approved power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used
- 2) Equipment is not to be used with flammable liquids.
- 3) Equipment is only to be installed and transported by manufacturer trained personnel. When the continuous flow rotor is used, the rotor with LUBRICATING UNIT shall be installed with the Ultracentrifuge by manufacturer trained personnel in the field.

Notes:

Products certified under Class C872106, C872186 have been certified under CSA's
ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca

