



Certificate of Compliance

Certificate: 70125009

Master Contract: 169502 (034862_C_000)

Project: 70125009

Date Issued: 2017-03-31

Issued to: Beckman Coulter Inc.
250 S. Kraemer Blvd
Brea, California 92822
USA

Attention: Bret Widdifield

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: *Shripal Pandya*
Shripal Pandya

PRODUCTS

CLASS - C872106 - LABORATORY EQUIPMENT-Electrical

CLASS - C872186 - ELECTRICAL EQUIPMENT FOR LABORATORY USE-Certified to US Standards

Automated Workstations,

Biomek i5, Models B87516, B87578, B87583 & B87584, Rated 100-240Vac, 50/60 Hz, 10 A

Biomek i7, Models B87579, B87580, B87581, B87582, B87585, B87586, B87587 & B87588, Rated 100-240Vac, 50/60 Hz, 10 A

Notes:

1. The above model is cord connected, Equipment Class 1, Pollution Degree 2, and Measurement Category II.
2. Mode of operation: Continuous
3. Environmental Conditions: Normal: 5 to 40 C, 2000 m max, relative humidity between 20% and 80% (non-condensing) at 30 degrees Celsius.



Certificate: 70125009

Project: 70125009

Master Contract: 169502

Date Issued: 2017-03-31

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-081:15

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

UL Standards:

UL Std. No. 61010-1 (3rd Edition)

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-081: 2015 (2nd Edition)

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

CONDITIONS OF ACCEPTABILITY

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) Equipment is only to be installed by manufacturer trained personnel.
- (3) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.

MARKINGS

The manufacturer is required to apply the following markings:

- Products shall be marked with the markings specified by the particular product standard.
- Products certified for Canada shall have all Caution and Warning markings in both English and French.

Additional bilingual markings not covered by the product standard(s) may be required by the Authorities Having Jurisdiction. It is the responsibility of the manufacturer to provide and apply these additional markings, where applicable, in accordance with the requirements of those authorities.



Supplement to Certificate of Compliance

Certificate: 70125009

Master Contract: 169502 (034862_C_000)

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70125009	2017-03-31	C/US Certification for Biomek i-Series 5 & 7 Automated Workstations