Beckman Coulter

MIAMI AND KENTUCKY REAGENTS MANUFACTURING

CERTIFICATE OF ANALYSIS

IDENTIFICATION:	Product Name:	Vi-CELL™FOCUS CONTROL
	Part Number:	175474
	Lot Number (Finished Product):	5474107F
	Lot Number (Sub-assembly):	5474107
	Date of Manufacture:	07-MAR-2023
	Expiration Date:	11-FEB-2025

TEST	SPECIFICATION	RESULTS	Pass/Fail
Final Particle Sizing	Modal size range must be between 1.818 to 2.22 microns.	Meets Specification	Pass

The product was manufactured in accordance with the US FDA Quality System Regulation (GMP).

APPROVAL	TITLE	DATE
[signature]	QUALITY ASSURANCE SCIENTIST I	30-mar-2023

Confidential - Company Proprietary