



Certificate of Analysis

Part Number	Description	Lot Number	Manufacture date	Expiration date	Storage Conditions
C09149	4M Vi-CELL Ctrl (4 x 10 ⁶ beads/mL)	210	2023-08-24	2024-08-24	2° to 8°C Do not freeze. Protect from light.

Test	Specification (x 10 ⁶ per mL)	Result (x 10 ⁶ per mL)	Pass/Fail
Concentration	3.80 – 4.20	3.97	PASS

Quality

Christy Majerczyk

Signature

15SEP2023

Date

Phone: 800-369-0333
Website: www.Beckman.com