DocuSign Envelope ID: 2C6E191A-42D6-400F-9DC2-085F8F45E620

Beckman Coulter

MIAMI AND KENTUCKY REAGENTS MANUFACTURING

CERTIFICATE OF ANALYSIS

IDENTIFICATION:	Product Name:	Vi-CELL™FOCUS CONTROL
	Part Number:	175474
	Lot Number (Finished Product):	5474113F
	Lot Number (Sub-assembly):	5474113
	Date of Manufacture:	22-NOV-2023
	Expiration Date:	04-NOV-2025

TEST	SPECIFICATION	RESULTS	Pass/Fail
Final Particle Sizing	Modal size range must be between 1.818 to 2.22 microns.	Meets Specification	Pass

The product was manufactured in accordance with the US FDA Quality System Regulation (GMP).

APPROVAL	TITLE	DATE
DocuSigned by: Esme (Esmeralda) Iglesias-Castillo	QUALITY ASSURANCE SCIENTIST I	01-Dec-2023

Signer Name: Esme (Esmeralda) Iglesias-Castillo
Signing Reason: I am the author of this document
Signing Time: 01-Dec-2023 | 8:05:11 AM EST
B683536A50444E1CBFAF6254E42D42C3

Confidential - Company Proprietary