DocuSign Envelope ID: 9EF1F4AB-CC2C-478A-8195-3EB93EE94724

# Beckman Coulter

*MIAMI AND KENTUCKY REAGENTS MANUFACTURING*

CERTIFICATE OF ANALYSIS

| | | |
|---|---|---|
| **IDENTIFICATION:** | Product Name: | Vi-CELL™FOCUS CONTROL |
| | Part Number: | 175474 |
| | Lot Number (Finished Product): | 5474114F |
| | Lot Number (Sub-assembly): | 5474114 |
| | Date of Manufacture: | 06-FEB-2024 |
| | Expiration Date: | 06-JAN-2026 |

| TEST | SPECIFICATION | RESULTS | Pass/Fail |
|---|---|---|---|
| Final Particle Sizing | Modal size range must be between 1.818 to 2.22 microns. | Meets Specification | Pass |

The product was manufactured in accordance with the US FDA Quality System Regulation (GMP).

DocuSigned by:

| APPROVAL | TITLE | DATE |
|---|---|---|
| Esme (Esmeralda) Iglesias-Castillo | QUALITY ASSURANCE SCIENTIST I | 12-Feb-2024 |

Signer Name: Esme (Esmeralda) Iglesias-Castillo
Signing Reason: I am the author of this document
Signing Time: 12-Feb-2024 | 12:56:40 PM EST

B683536A50444E1CBFAF6254E42D42C3



Confidential - Company Proprietary